

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0508D

System: GNDRAD PSP

Site: Darwin

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PSP	31284	Gndrad	9/7/2004	7.9318	na	8/9/2005	23:00	23:30

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PSP	31283	Gndrad	8/9/2005	8.393	na	X

Replacement completed by: Gary Eckert, Mike Alsop, Bill
Porch

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: